

Device Information

\*\*\*\*\*

Device: tempmate-S1

Version: V2.0

Model No. : Customized

Serial Number: S116060097-14

Internal ID: TEST

Description: tempmate-S1 V2 Temperature Recording

Log Interval: 00H 01M 00S

Start Mode: Start by timing

Start Timing: 01-Jul-16 08:30:00

Stop Mode: Stop by button

Note:

All temperatures are in Degree Celsius.

All times shown are based on UTC +02:00 and 24-Hour clock [DD-MMM-YYYY HH:MM:SS]

Alarm Configuration/Summary

\*\*\*\*\*

Alarm At:	01-Jul-16 09:50:00		
Alarm Zone	Alarm Delay	Total Time	No. of Violation
H3: over: 35.0	00D 00H 00M 00S(Sin)	00D 02H 05M 00S	3
H2: over: 30.0	00D 00H 00M 00S(Sin)	00D 02H 22M 00S	1
H1: over: 25.0	00D 00H 00M 00S(Sin)	00D 02H 25M 00S	1
L1: below: 15.0	00D 00H 00M 00S(Sin)	00D 03H 32M 00S	2
L2: below: 10.0	00D 00H 00M 00S(Sin)	00D 03H 19M 00S	2

Logging Summary

\*\*\*\*\*

Highest Temperature: 41.9

Lowest Temperature: -23.5

Average Temperature: 15.4

MKT: 28.1

Start Time: 01-Jul-16 08:30:00

Stop Time: 01-Jul-16 15:50:00

Elapsed Time (stop by button) : 00D 07H 20M 00S

Data Points: 441

Marked Events

\*\*\*\*\*

1: 01-Jul-16 09:46:00 21.3

2: 01-Jul-16 11:56:00 13.9

3: 01-Jul-16 14:24:00 28.6

Recorded Temperatures

\*\*\*\*\*

Date	Time	Temperature
01-Jul-16	08:30:00	24.6
01-Jul-16	08:31:00	24.6
01-Jul-16	08:32:00	24.6
01-Jul-16	08:33:00	24.6
01-Jul-16	08:34:00	24.6
01-Jul-16	08:35:00	24.6
01-Jul-16	08:36:00	24.6
01-Jul-16	08:37:00	24.6
01-Jul-16	08:38:00	24.6
01-Jul-16	08:39:00	24.6
01-Jul-16	08:40:00	24.6
01-Jul-16	08:41:00	24.6
01-Jul-16	08:42:00	24.5
01-Jul-16	08:43:00	24.5
01-Jul-16	08:44:00	24.5
01-Jul-16	08:45:00	24.6
01-Jul-16	08:46:00	24.6
01-Jul-16	08:47:00	24.6