



# Data Report

File Created Date: 19-Jul-23 13:26:54

Note: All times shown are based on UTC +02:00 and 24-Hour clock [DD-MMM-YY HH:MM:SS]

## Device Information

Serial Number: XS23060001-23	Log Interval: 00H 00M 10S	Temperature Type: °C
Model Number: Customized	Start Mode: Manual	
Version: V0.6	Start Delay: 00D 10H 00M 00S	

## Trip Information

Internal ID: 0000001
Description: Temperature recording. first testing

## Logging Summary

Highest Temperature: 27.9 °C	Record Mode: Stop when full	Start Time: 19-Jul-23 02:34:07
Lowest Temperature: 23.2 °C	Stop Mode: Stop by Button	Stop Time: 19-Jul-23 13:26:47
Average Temperature: 25.3 °C	Data Points: 3917	Elapsed Time: 00D 10H 52M 40S
MKT: 25.3 °C	Alarm At: 19-Jul-23 08:11:27	

## Marked Events

N/A

Alarm Zone	Alarm Delay	Total Time	No. of Violations	Status
H1: over 25.0 °C	00D 00H 20M 00S(Sin)	00D 04H 44M 10S	3	Alarm
SZ: 25.0 to 9.0 °C	unlimited	00D 06H 08M 30S		
L1: below 9.0 °C	00D 00H 00M 00S(Sin)	00D 00H 00M 00S	0	OK

## Temperature [°C]

Normal: — Mark: — No Data: .....

